

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
13 Parameter (Alkalinity, Ca, Cl, NO <sub>3</sub> , FI, Fe, Hardness, Mn, Na, pH, SO <sub>4</sub> , TDS, Special QT)	<a href="#">See Individual test below</a>	<a href="#">See Individual test below</a>	\$ 273.00	(1) 500ml P (1) 250 ml P (1) 125ml P H <sub>2</sub> SO <sub>4</sub> (1) 125ml P HNO <sub>3</sub> (1) Idexx Bacti	Cool 4C	<a href="#">See Individual test below</a>	14 calendar days	<a href="#">See Individual test below</a>
Langlier Index (Alkalinity, Hardness, pH, TDS)	<a href="#">See Individual test below</a>	<a href="#">See Individual test below</a>	\$ 94.00	(1) 125ml P HNO <sub>3</sub> (2) 500ml P	Cool 4C	<a href="#">See Individual test below</a>		<a href="#">See Individual test below</a>
Phase II (Ba, Cd, Cr, FI, Hg, Se)	<a href="#">See Individual test below</a>	<a href="#">See Individual test below</a>	\$ 135.00	(1) 250ml P (1) 125ml P	250 ml preserved in lab, HNO <sub>3</sub> <2	<a href="#">See Individual test below</a>	21 days	<a href="#">See Individual test below</a>
Phase V (Antimony, Be, CN, Ni, SO <sub>4</sub> , Thallium)	<a href="#">See Individual test below</a>	<a href="#">See Individual test below</a>	\$ 148.00	(2) 125 ml P (1) 125ml P NaOH	Cool 4C	<a href="#">See Individual test below</a>	21 days	<a href="#">See Individual test below</a>
Total Water Quality (New Well)	<a href="#">See Individual test below</a>	<a href="#">See Individual test below</a>	\$3,472.00			Varies for each kit	8 wks	<a href="#">See Individual test below</a>
Lead/Copper	<a href="#">See Individual test below</a>	EPA 200.8 Rev 5.4	\$ 38.00	1L P wide-mouth	Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	<a href="#">See Individual test below</a>
Asbestos *	0.20 MFL	EPA 100.2	\$ 160.00	1L P	Cool 4C	48 hours	6 wks	7 MFL
Alkalinity as CaCO <sub>3</sub>	20 mg/L	SM 2320B	\$ 20.00	125 ml P	Cool 4C	14 days	7 days	-
Aluminum	2 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	
Ammonia (by distillation)	0.05 mg/L	EPA 350.1 Rev 2.0	\$ 30.00	250 ml P	Cool 4C, H <sub>2</sub> SO <sub>4</sub> <2	28 days	14 days	-
Ammonia (by diffusion)	0.05 mg/L	EPA 350.1 Rev 2.0	\$ 22.00	250 ml P	Cool 4C, H <sub>2</sub> SO <sub>4</sub> <2	28 days	14 days	-
Antimony	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	6 µg/L
AOC (Assimilable Organic Carbon (sent to UL)**)	10 ug ACE/L	SM9217B	\$ 330.00	500ml A,G Collect M-W	Cool 4C	72 hrs		
Arsenic	1 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	10 µg/L
Barium	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	2000 µg/L

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
Beryllium	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	4 µg/L
Bicarbonate/ Carbonate *	5 mg/L	SM2320 B	\$ 13.00	250 ml P	Cool 4C	14 days	21 days	
Bismuth *	0.1 mg/L	EPA 200.8 Rev 5.4	\$ 28.00	500 ml P	HNO <sub>3</sub> <2	6 months	21 days	
BOD/CBOD	2 mg/L	EPA 405.1 / SM 5210-B	\$ 54.00	1L cubie	Cool 4C	48 hours	7 days	-
Boron *	0.1 mg/L	EPA 200.7	\$ 19.00	500 ml P	pH<2 HNO <sub>3</sub>		14 days	
Bromate *	1 ug/L	EPA317.0,Rev 2.0	\$ 166.00	125 ml P	1% EDA, Cool 4C	28 days	21 days	.010 mg/L
Cadmium	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	5 µg/L
Calcium, Envir. & D.W.	0.15 mg/L	SM 3111B	\$ 21.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	14 days	-
Chloride	1 mg/L	Lachat 10-117-07-1-A	\$ 21.00	125 ml P	None required	28 days	14 days	-250 mg/L 2°MCL
Chlorite * (as of 9/4/15)	10 ug/L	EPA 300.0 Rev 2.1	\$ 37.00	125 ml P	Cool 4C, EDA	14 days	21 days	1 mg/L
Chromium	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	100 µg/L
Cobalt	0.5 ug/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	
COD	12 mg/L	Hach8000	\$ 52.00	125 ml P	Cool 4C, H <sub>2</sub> SO <sub>4</sub> <2	28 days	14 days	-
Colilert (Total Coliform and E.coli)	0 cfu/100 ml	SM 9223B	\$ 15.00	Idexx bacti	Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub> (and <10C for NPDES)	30 hours (8hrs for NPDES)	3 days	-
Colilert Quantitray (Total Coliform and E.coli)	0 MPN/100 ml	SM 9223B	\$ 17.00	Idexx bacti	Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub> (and <10C for NPDES)	30 hours (8hrs for NPDES)	3 days	-
Colisure (Total Coliform and E.coli)	0 cfu/100 ml	SM 9223B	\$15	Idexx Bacti	Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub>	30 hours	3 days	
Colisure QT(Total Coliform and E.coli)	0 MPN/100 ml	SM 9223B	\$17	Idexx Bacti	Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub>	30 hours	3 days	
Conductivity	1.0 umho/cm	SM 2510B	\$ 15.00	125 ml P	Cool 4C	28 days	7 days	-

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
Copper	2.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	1300 µg/L Action Level
Cryptosporidium/Giardia**	<.093/L	EPA 1623	\$ 472.00	10 L cubie or field filter	Cool 4C	96 hrs		
Cyanide, Free (regulated)	0.05 mg/L	SM4500CN-F	\$ 47.00	125 ml P	Cool 4C, NaOH>12	14 days	14 days	200 µg/L
Cyanide, Total*	.01 mg/L	SM4500CN-E	\$ 33.00	250 ml P	Cool 4C, NaOH>12	14 days	14 days	
DIC (dissolved inorganic carbon)*		SW9060	\$ 50.00	500ml A, G	Cool 4C	28 days	2 weeks	
DOC (dissolved organic carbon)*	1 mg/L	SM 5310C	\$ 41.00	(2)40ml amber vials	Cool 4C, H <sub>2</sub> SO <sub>4</sub> <2	28 days	2 weeks	
Dissolved Oxygen	1 mg/L	SM4500-O-G	\$ 15.00	1L cubie	Cool 4C			
Fecal Coliform by Colilert 18QT	0 MPN/100 ml	SM 9223B QT	\$ 17.00	Idexx bacti	Cool <10C	8 hrs for NPDES 30hr for DW 24hr-for non-compliance	4 days	-
Fluoride(EPA9214 in LIMS)	0.20 mg/L	Lachat 10-109-12-2-A	\$ 18.00	125 ml P	None required	28 days	7 days	4 mg/L
Heterotrophic Plate Count	0 MPN/ml 2 col/1ml	SM 9215E	\$ 39.00	Idexx bacti	Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub> , Cool <10C	8 hr DW 24 hr non-DW If >2hr to deliver, keep <10C	4 days	-
Hydroxide as OH*	1 mg/L	A2320B	\$ 13.00					
Iron Related Bacteria	1 cfu/100	IRB-BART	\$ 65.00	Idexx bacti	Cool <10C or recd w/in 4hrs	24 hours	14 days	
Iron, Envir. & D.W.	75 µg/L	SM 3111B	\$ 21.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	14 days	300 µg/L 2°MCL
Lead	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	15 µg/L Action Level
Legionella pneumophila	0 MPN/100 ml	Idexx Legiolert ASTM D8429-1	\$ 65.00	Idexx bacti	Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub> Insulated cooler to maintain ambient temp	48 hrs	7 days	

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
Magnesium - Envir. & D.W.	0.15 mg/L	SM 3111B	\$ 21.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	14 days	-
Manganese	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	0.05 mg/L 2°MCL
Mercury, Envir. & D.W.	0.2 µg/L	EPA 245.1 Rev 3.0	\$ 41.00		Preserved in lab, HNO <sub>3</sub> <2	28 days	21 days	2 µg/L
Microwave Digestion	NA	SM 3015	\$ 16.00			6 months	NA	-
Molybdenum	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	
Nickel	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	
Nitrate/Nitrite	0.05 mg/L	Lachat 10-107-04-1-A	\$ 16.00	125 ml P	Cool 4C recommended, H <sub>2</sub> SO <sub>4</sub> <2	28 days	7 days	10 mg/L
Nitrogen-Nitrite	0.1 mg/L	Lachat 10-107-05-1-A	\$ 32.00	125 ml P	Cool 4C	48 hours	7 days	1 mg/L
pH	NA	EPA 150.1 / SM4500-H <sup>+</sup> B	\$ 20.00	125 ml P	None required	0.25 hours	7 days	-
Phosphate-Ortho, as P	0.02 mg/L	EPA 365.1 Rev 1.0	\$ 23.00	125 ml P	Filter, cool 4C	48 hours	7 days	-
Phosphorus-Total	0.04 mg/L	EPA 365.4	\$ 30.00		Cool 4c, H <sub>2</sub> SO <sub>4</sub> <2	28 days	14 days	-
Potassium - Envir. & D.W.	0.30 mg/L	SM 3111B	\$ 21.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	14 days	-
Residual Chlorine	0.1 mg/L	SM4500-CL-G	\$ 26.00	125 ml P	None required	6 hours	7 days	-
Salinity *	0.1	SM2520B	\$ 14.00	250ml P	Cool 4C	28 days		
Selenium	2.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	50 µg/L
Silica *	0.1 mg/L	E200.7	\$ 28.00	500 ml P	HNO <sub>3</sub> <2	180 days	14 days	
Silicon * (Si)	0.05 mg/L	E200.7	\$ 28.00	500 ml P	HNO <sub>3</sub> <2	180 days	14 days	
Silver	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	100 ug/L 2°MCL
Sodium - Envir. & D.W.	0.15 mg/L	SM 3111B	\$ 21.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	14 days	

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
Strontium	1 ug/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	
Sulfate	10 mg/L	SM 4500 SO <sub>4</sub> -E	\$ 25.00	125ml P	Cool 4C	28 days	14 days	-250 mg/L 2°MCL
Sulfide *	0.5 mg/L	A4500-S F	\$ 45.00	1L P	Cool 4C, zinc acetate, NaOH>9	7 days	14 days	
Sulfate Reducing Bacteria	1 cfu/100	SRB-BART	\$ 65.00	Idexx bacti	Cool <10C or set up w/in 4hrs	24 hours	14 days	
Sulfur, Total *	0.1 mg/L	SW6010B EPA 200.7	\$ 45.00 \$90 if dig.	250 ml P	pH<2 HNO <sub>3</sub>	6 months	14 days	
SSC (Suspended Sediment Concentration)	5 mg/L	ASTM D3977-97	\$ 32.00	8 oz clear G	None required	28 days	21 days	-
Surfactants *	1 mg/L	A5540C	\$ 13.00	500 ml P	Cool 4C	48 hours	14 days	
Thallium	0.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	2 µg/L
Titanium *	0.1 mg/L	EPA 200.8 Rev 5.4	\$ 28.00	500 ml P	HNO <sub>3</sub> <2	6 months	21 days	
TOC *	0.5 mg/L	SM5310-B	\$ 41.00	(2) 40ml AG vials	Cool 4C, H <sub>2</sub> SO <sub>4</sub> <2	28 days	21 days	
Total Dissolved Solids	10 mg/L	SM 2540-C	\$ 29.00	500 ml P	Cool 4C	7 days	7days	-500 mg/L 2°MCL
Total Hardness	4 mg/l	SM 2340C	\$ 25.00	125 ml P	HNO <sub>3</sub> <2	6 months	7 days	-
Total Kjeldahl Nitrogen	0.50 mg/l	EPA 351.2 Rev 2.0	\$ 30.00		Cool 4c, H <sub>2</sub> SO <sub>4</sub> <2	28 days	14 days	-
Total Suspended Solids	5 mg/l	EPA 160.2	\$ 21.00	500 ml P	Cool 4C	7 days	7 days	-
Turbidity	0.1 NTU	EPA 180.1 Rev 2.0	\$ 12.00	125 ml P	None required	48 hrs	7 days	1 NTU only for surface water
Uranium (Mass)	0.5 ug/L	EPA 200.8 Rev 5.4	\$ 19.00	125 ml P	Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	-30 ug/L
Vanadium	0.5 ug/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
Zinc	2.5 µg/L	EPA 200.8 Rev 5.4	\$ 19.00		Preserved in lab, HNO <sub>3</sub> <2	6 months	21 days	-5 mg/L 2°MCL
1664 * Non-Polar Material (SGT-HEM = TPH) Oil & Grease (HEM = O&G)	5 mg/l	EPA 1664	\$ 153.00	(2) 1 L AG	Cool 4C, H <sub>2</sub> SO <sub>4</sub> or HCl <2	28 days	28 days	HEM 10 mg/L
Diesel Range Organics (DRO) * TPH as diesel	1 mg/L	OA-2	\$ 98.00	1 L AG	Cool 4C	7 days Collection to extraction; 40 days Extraction to analysis	28 days	
Gasoline Range Organics (GRO) *	1 mg/L	OA-1	\$ 98.00	(3) 40ml G vials	Cool 4C, HCl<2	14 days	28 days	
Gross Alpha *	MDC	EPA 900.0	*\$ 55.00	500ml P	HNO <sub>3</sub> <2	6 months	6 wks	-15 pCi/L
Radon in Water (EPA 913.0 in LIMS)	15 pCi/L	SM 7500-RN	\$ 110.00	(4) 40ml G clear vials	None required	4 days from collect date/time	7 days	-300 pCi/L
Radium-226 * SM7500-Ra for the calc.	MDC	EPA 903.1	*\$ 83.00	1L P	HNO <sub>3</sub> <2	6 months	6 wks	-5 pCi/L for combined
Radium-228 * SM7500-Ra for the calc.	MDC	EPA 904.0	*\$ 83.00	1L P	HNO <sub>3</sub> <2	6 months	6 wks	-5 pCi/L for combined
Thorium, Isotopic*	2 pCi/L	EPA 907.0	\$ 141.00	2 L P	HNO <sub>3</sub> <2, Keep 4C if not pres. in field	6 months		
Uranium, Isotopic*	1 pCi/L	HASL300 U02	*\$ 130.00	500ml P	HNO <sub>3</sub> <2	6 months	6 wks	-20 pCi/L
1,2 Dibromoethane (EDB)	0.01 µg/L	EPA 504.1 Rev 1.1	\$ 157.00	(2) 40 ml clear G vials + fblk	Cool 4C, Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub>	14 days Collection to extraction; 24 hours Extraction to analysis	17 days	0.05 µg/L
1,2 Dibromo-3-Chloropropane (DBCP)	0.01 µg/L							-0.2 µg/L
PCB-1016 Aroclor	0.1 µg/L	EPA 505 Rev 2.1	\$ 159.00	(2) 40 ml clear G vials	Cool 4C, Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub>	14 days Collection to extraction;	17 days	-Total 0.5µg/L
PCB-1221 Aroclor	2 µg/L							-
PCB-1232 Aroclor	0.5 µg/L					24 hours		-
PCB-1242 Aroclor	0.3 µg/L					Extraction to		-

Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)		
PCB-1248 Aroclor	0.2 µg/L					analysis		-		
PCB-1254 Aroclor	0.1 µg/L							-		
PCB-1260 Aroclor	0.2 µg/L							-		
Toxaphene	1 µg/L							3 µg/L		
2,4-D	1 µg/L	EPA 515.3 Rev 1.0	\$ 258.00	(2) 60 ml AV	Cool 4C	14 days Collection to extraction;	21 days	70 µg/L		
2,4,5-TP (Silvex)	0.5 µg/L				NC – none				50 µg/L	
Dalapon	1 µg/L				C – Na2S2O4	14 days Extraction to analysis		200 µg/L		
Dicamba	1.0 µg/L								-	
Dinoseb	2.0 µg/L								7 µg/L	
Pentachlorophenol	0.2 µg/L								1 µg/L	
Picloram (Tordon)	3.0 µg/L								500 µg/L	
Bromodichloromethane (THM)	0.5 µg/L	EPA 524.2 Rev 4.1	\$ 237.00	(2) 40 ml G vials + fblk	Cool 4C	14 days	14 days	-**TTHM is 80 µg/L		
Bromoform (THM)	0.5 µg/L				Qtrly – 3 + fblk			NC – HCl<2		-
Dibromochloromethane (THM)	0.5 µg/L							C – ascorbic acid		-
1,1,1,2-Tetrachloroethane	0.5 µg/L								-	
1,1,1,-Trichloroethane	0.5 µg/L								200 µg/L	
1,1,2,2-Tetrachloroethane	0.5 µg/L								-	
1,1,2-Trichloroethane	0.5 µg/L								5 µg/L	
1,1-Dichloroethane	0.5 µg/L								-	
1,1-Dichloroethene	0.5 µg/L								7 µg/L	
1,1-Dichloropropene	0.5 µg/L								-	
1,2,3-Trichloropropane	0.5 µg/L								-	
1,2,4-Trichlorobenzene	0.5 µg/L								70 µg/L	
1,2,4-Trimethylbenzene	0.5 µg/L								-	
1,2,3-Trichlorobenzene	0.5 µg/L								-	

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
1,2-Dibromo-3-Chloropropane (DBCP)	0.5 µg/L							-
1,2-Dichlorobenzene	0.5 µg/L							600 µg/L
1,2-Dichloroethane	0.5 µg/L							5 µg/L
1,2-Dichloropropane	0.5 µg/L							5 µg/L
1,3-Dichloropropane	0.5 µg/L							-
1,3,5-Trimethylbenzene	0.5 µg/L							-
1,3-Dichlorobenzene	0.5 µg/L							-
1,4-Dichlorobenzene	0.5 µg/L							75 µg/L
2,2-Dichloropropane	0.5 µg/L							-
2-Chlorotoluene	0.5 µg/L							-
4-Chlorotoluene	0.5 µg/L							-
Benzene	0.5 µg/L							5 µg/L
Bromobenzene	0.5 µg/L							-
Bromochloromethane	0.5 µg/L							-
1,2-Dibromoethane (EDB)	0.5 µg/L							-
Bromomethane	0.5 µg/L							-
Carbon Tetrachloride	0.5 µg/L							5 µg/L
Chlorobenzene	0.5 µg/L							100 µg/L
Chloroethane	0.5 µg/L							-
Chloroform (THM)	0.5 µg/L							-
Chloromethane	0.5 µg/L							-
Cis-1,2-dichloroethene	0.5 µg/L							70 µg/L
Cis-1,3-dichloropropene	0.5 µg/L							-
Dibromomethane	0.5 µg/L							-
Dichlorodifluoromethane	0.5 µg/L							-



## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
Dichloromethane (methylene chloride)	0.5 µg/L							5 µg/L
Ethylbenzene	0.5 µg/L							700 µg/L
Hexachlorobutadiene	0.5 µg/L							-
Isopropylbenzene	0.5 µg/L							-
M,P-Xylenes	0.5 µg/L							Total Xylenes 10,000 µg/L
O-Xylene	0.5 µg/L							-
Methyl-T-butyl ether (MTBE)	0.5 µg/L							-
Naphthalene	0.5 µg/L							-
n-Butylbenzene	0.5 µg/L							-
n-Propylbenzene	0.5 µg/L							-
P-Isopropyltoluene	0.5 µg/L							-
Sec-butylbenzene	0.5 µg/L							-
Styrene	0.5 µg/L							100 µg/L
Tert-butylbenzene	0.5 µg/L							-
Tetrachloroethene	0.5 µg/L							5 µg/L
Toluene	0.5 µg/L							1000 µg/L
Trans-1,2-dichloroethene	0.5 µg/L							100 µg/L
Trans-1,3-dichloropropene	0.5 µg/L							-
Trichloroethene	0.5 µg/L							5 µg/L
Trichlorofluoromethane	0.5 µg/L							-
Vinyl Chloride	0.5 µg/L							2 µg/L
THM's								80 µg/L
Alachlor	0.1 µg/L	EPA 525.2 Rev 2.0	\$ 247.00	(2) 1L AG	Cool 4C	14 days to Extraction;	21 days	2 µg/L
Aldrin	0.1 µg/L				NC – HCl<2			-
Atrazine	0.08 µg/L				C – sodium sulfite,	30 days		3 µg/L

Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)	
					HCl in lab to <2	Extraction to Analysis			
Benzo(a)pyrene	0.06 µg/L							0.2 µg/L	
Butachlor	0.1 µg/L							-	
Butylate	0.1 µg/L							-	
Chlorpyrifos (=Dursban=Lorsban)	0.1 µg/L							-	
Cyanazine (=Bladex)	0.15 µg/L							-	
Di(2-ethylhexyl)phthalate	2 µg/L							6 µg/L	
Di(2-ethylhexyl)adipate	0.5 µg/L							400 µg/L	
Dieldrin	0.1 µg/L							-	
Endrin	0.1 µg/L							2 µg/L	
Fonofos (=Dyfonate)	0.1 µg/L							-	
Heptachlor	0.08 µg/L							0.4 µg/L	
Heptachlor Epoxide	0.04 µg/L							0.2 µg/L	
Hexachlorobenzene	0.1 µg/L							1 µg/L	
Hexachlorocyclopentadiene	0.1 µg/L							50 µg/L	
Lindane	0.04 µg/L							0.2 µg/L	
Methoxychlor	0.1 µg/L							40 µg/L	
Metolachlor	0.1 µg/L							-	
Metribuzin	0.5 µg/L							-	
Propachlor	0.1 µg/L							-	
Simazine	0.1 µg/L					-4 µg/L			
Total Chlordane	0.3 µg/L					2 µg/L			
Trifluralin (=Treflan)	0.1 µg/L					-			
3-Hydroxy carbofuran	0.5 µg/L	EPA 531.1 Rev 3.1	\$ 223.00	(2) 40ml G vials	Cool 4C	28 days	21 days	-	
Aldicarb	0.5 µg/L				NC – ChlorAC			Any 2 analytes not	-3 ug/L
Aldicarb sulfone	0.5 µg/L				C – ChlorAC +				-2 ug/L

## Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
					Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub>		to exceed 7 ug/L	
Aldicarb sulfoxide	0.5 µg/L							-4 ug/L
Carbaryl (Sevin)	0.5 µg/L							-
Carbofuran	0.5 µg/L							40 µg/L
Methomyl	0.5 µg/L							-
Oxamyl	0.5 µg/L							200 µg/L
Glyphosate (=Roundup)	5 µg/L	EPA 547	\$ 216.00	(2) 40 ml G vials + fblk	Cool 4C C-4mg Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub> , Cool 4C	14 days 18 months frozen	21 days	700 µg/L
Endothall *	9 µg/L	EPA 548.1 Rev 1.0	*\$ 160.00	(2) 40ml AV	Cool 4C, Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub>	7 days collection to Extraction; 14 days Extraction to Analysis	6 wks	-100 µg/L
Diquat	1.0 µg/L	EPA 549.2 Rev 1.0	\$ 241.00	1L brown P	Cool 4C, Na <sub>2</sub> S <sub>2</sub> O <sub>4</sub>	7 days to Extract; 21 days Analysis	21 days	20 µg/L
Paraquat	1.0 µg/L							-
BCAA (not regulated as part of HAA5)	1.0 µg/L	EPA 552.2 Rev 1.0	\$ 248.00	(2) 60 ml AV	Cool 4C, NH <sub>4</sub> Cl	14 days Collection to Extraction;	21 days	60 µg/L
DBAA	0.5 µg/L							
DCAA	1.0 µg/L					7 days Extraction to Analysis		
MBAA	1.0 µg/L							
MCAA	2.0 µg/L							
TCAA	0.5 µg/L							
Dioxin *	4.8e-06 µg/L = .005 ng/L	EPA 1613	\$ 290.00	1L AG	Cool 4C	30 days	6 wks	0.00003 µg/L Or .03 ng/L
Ethylene Glycol*	10 mg/L	EPA 8015B	\$ 98.00	(1) VOA vial, no pres.	Cool 4C	14 days		
8260B Regular List*		EPA 8260B	\$ 169.00	(3) VOC vials	Cool 4C, HCl	14 days		
8260B, Long List*		EPA 8260B	\$ 245.00	(3) VOC vials	Cool 4C, HCl	14 days		

Nebraska DHHS Laboratory

(pricing effective 7/1/13 – 12/31/23)

Analyte	Report Level (RL)	Method	Price	Sample Size/ Container	Preservative	Holding Time	Turn Around Time	Maximum Contaminant Level (MCL)
8260B, BTEX*		EPA 8260B	\$ 117.00	(3) VOC vials	Cool 4C, HCl	14 days		
EPA 624*		EPA 624	\$ 130.00	(3) VOC vials	Cool 4C, HCl	10 days		
EPA 625*		EPA 625	\$ 370.00	(2) 1L AG	Cool 4C			
PCB's in Oil*		SW8082	\$ 120.00	40ml VOA vial	Cool 4C			
PCB's in liquid*		SW8082	\$	1L AG	Cool 4C			
MPA			\$425			96 hrs	10 days	
EPA 8330B (Explosives) (17 compounds)	1.0 ug/L	EPA 8330B	\$245	(2) 250ml AG	Cool 4C	7d to Extract 40d Extract to Analysis		
Hexavalent Chromium	10 ug/L	A3500CR-B	\$65	500ml P	Cool 4C	24 hr (notify lab when coming)		

**Container Types**

		<u>Preservatives</u>	
<b>P</b>	<b>Plastic (HDPE)</b>	<b>NaOH</b>	<b>Sodium Hydroxide</b>
<b>AG</b>	<b>Amber Glass</b>	<b>HCl</b>	<b>Hydrochloric Acid</b>
<b>G</b>	<b>Glass</b>	<b>HNO3</b>	<b>Nitric Acid</b>
<b>AV</b>	<b>Amber Vials</b>	<b>H2SO4</b>	<b>Sulfuric Acid</b>
		<b>Na2S2O4</b>	<b>Sodium Thiosulfate</b>
		<b>NH4Cl</b>	<b>Ammonium Chloride</b>
		<b>H3PO4</b>	<b>Phosphoric Acid</b>

\* = Pace Analytical or their subcontract lab

C = Chlorinated sites

NC = Non-chlorinated sites

NOTE: Last changed 5/1/23